



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Engineering Analytics Laboratories
1299 Goode Dr. NE, Palm Bay, FL 32907

*(Hereinafter called the Organization) and hereby declares that Organization is accredited
in accordance with the recognized International Standard:*

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the
operation of a laboratory quality management system
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Chemical and Mechanical Testing
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

December 27, 2023

Issue Date:

December 27, 2023

Expiration Date:

March 31, 2026

Accreditation No.:

122465

Certificate No.:

L23-946

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a
continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjilabs.com*



Certificate of Accreditation: Supplement

Engineering Analytics Laboratories

1299 Goode Dr. NE, Palm Bay, FL 32907

Contact Name: Mr. Kenneth Bukowski Phone: 321-720-6578

Accreditation is granted to the facility to perform the following testing:

FLEX CODE	FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED
F1, F2 or F5	Chemical Mechanical ^F	Inorganic and organic materials	Melting Point	ASTM D967	DSC
F1, F2 or F5			Curie Point/Weight Change	ASTM E1582	TGA
F1, F2 or F5			Structural Transition	ASTM D1867 (DMA); ASTM D967 (DSC)	DMA, DSC

- The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location.
- Flex Code:
 - F1-Introduction of the testing of a new item, material, matrix, or product for an accredited test method
 - F2-Introduction of a new version of an accredited standard method (with no modifications)
 - F3-Introduction of a new parameter/component/analyte to an accredited test method
 - F4- Introduction of a new version or modifications of an accredited non-standard method
 - F5-Introduction of a new method that is equivalent to an accredited method (using same technology or technique)